HOHENSTEIN



PPprint GmbH Gottlieb-Keim-Str. 60 - 95448 Bayreuth - Germany

Sample P-filament 721 natural Polypropylen

According to the report no. **20.8.5.1383** dated on 19 October 2020 we confirm that the above mentioned item was tested at the Hohenstein Laboratories GmbH & Co. KG and is awarded with the certificate confirming "Biological Safety".

The following listed biocompatibility test was carried out according to the appropriate standards and was successfully passed:

• the Cytotoxicity Test according to DIN EN ISO 10993-5

The certificate is valid until 31 October 2021

hino famme

Dr. Timo Hammer Chief Executive Officer



Hohenstein Laboratories GmbH & Co. KG | Schlosssteige 1 | 74357 Bönnigheim | Germany